

# BIOCOMPATIBILITY CERTIFICATE

**Testmaterial:** Sensor hand piece assembly incl. sensor tube

**Supplier:** ndd Medizintechnik AG, Technoparkstrasse 1, 8005, Zürich,  
Switzerland

**Studies performed:** The following studies were performed in order to determine the biocompatibility of the product according to ISO 10993-1:


**CYTOTOXICITY  
SKIN IRRITATION  
LOCAL LYMPH NODE ASSAY**

**Results:** The product did not show any adverse effects in the studies performed. Therefore, the biocompatibility of the test material was proved.

**BSL BIOSERVICE Scientific Laboratories GmbH Munich**

Behringstraße 6

D-82152 Planegg



Dr. Varun Ahuja

Biological Safety Testing

Date: May 6, 2010



**BSL BIOSERVICE** Scientific Laboratories GmbH

Behringstrasse 6/8 · 82152 Planegg, Germany  
Telefon +49(0)89-899 65 00 · Fax +49(0)89-899 65 011  
e-mail: info@bioservice.com · www.bioservice.com  
Geschäftsführer: Dr. Wolfram Riedel  
Amtsgericht München, HRB 109 770  
Erfüllung und Gerichtsstand München  
NORD/LB Norddeutsche Girozentrale, BLZ 250 500 00, Kto. 151 328 523, Swift-BIC: NOLADE2HXXX, IBAN: DE76 2505 0000 0151 3285 23  
Deutsche Bank München, BLZ 700 700 24, Kto. 9 407 750, Swift-BIC: DEUTDE33MUC, IBAN: DE52 7007 0024 0940 7750 00



Akkreditiert durch  
Zentralstelle der Länder  
für Gesundheitsschutz  
bei Arzneimitteln  
und Medizinprodukten  
ZLG-P-986.96.01